

Hearing Aid Compatibility Rules - Update

TCB Workshop
October 2016

Jim Szeliga
Laboratory Division
Office of Engineering and Technology
Federal Communications Commission




HAC Update

- Update - Coming in just over a year (2018).
 - Revised scope (Fourth Report and Order and proposed Rulemaking - 15-155) changes HAC in that after January 2018---the required revised scope covers:
 - Wi-Fi calling
 - Wi-Fi expanded service
 - Wireless Non CMRS Voice services:
 - » Pre-installed by manufacturer
 - » Installed by carrier/authored by manufacturer prior to sale.
 - » Post sale if offered by manufacturer/carrier



HAC Update

- Plan to modify KDB Publication 285076 20.19, Hearing Aid Compatibility- HAC ASAP - first as a draft for comment.
- Test T-Coil – test can be done now.
 - Clarify the transition to 2018
 - For testing T-Coil test codec at defined reference of average talker level or
 - Our current plan:
 - If reference level is undefined use **-20 dBm0**.



Questions and Answers

Thanks!